

Miniaturized Gas Correlation Radiometer for the Detection of Trace Gases in the Martian Atmosphere

Authors: Hilary R. Melroy (GWU/GSFC, Code 694), Emily L. Wilson (NASA/GSFC, Code 694), Elena Georgieva (UMBC/GSFC, Code 614.0)

We present a miniaturized and simplified version of a gas correlation radiometer (GCR) capable of simultaneously mapping multiple trace gases and identifying active regions on the Mars surface.¹ Gas correlation radiometry (GCR) has been shown to be a sensitive and versatile method for detecting trace gases in Earth's atmosphere.^{2,3} Reduction of the size and mass of the GCR was achieved by implementing compact, light-weight 1 mm inner diameter hollow-core optical fibers (hollow waveguides) as the gas correlation cells. In a comparison with an Earth orbiting CO₂ GCR instrument, exchanging the 10 m multipass cells with hollow waveguide gas correlation cells of equivalent pathlength reduces the mass from ~150 kg to ~0.5 kg, and reduces the volume from 1.9 m x 1.3 m x 0.86 m to a small bundle of fiber coils approximately 1 meter in diameter by 0.05 m in height (mass and volume reductions of >99%)^{4,5}.

A unique feature of this instrument is its stackable module design, with a single module for each trace gas. Each of the modules is self-contained, and fundamentally identical; differing by the bandpass filter wavelength range and gas mixtures inside the hollow-waveguide absorption cells. The current configuration contains four stacked modules for simultaneous measurements of methane (CH₄), formaldehyde (H₂CO), water vapor (H₂O), and deuterated water vapor (HDO) but could easily be expanded to include measurements of additional species of interest including nitrous oxide (N₂O), hydrogen sulfide (H₂S), methanol (CH₃OH), and sulfur dioxide (SO₂), as well as carbon dioxide (CO₂) for a simultaneous measure of mass balance.

Preliminary results indicate that a 1 ppb detection limit is possible for both formaldehyde and methane with one second of averaging. Using non-optimized components, we have demonstrated an instrument sensitivity equivalent to ~30 ppb for formaldehyde, and ~500 ppb for methane. We expect custom bandpass filters and 6 m long waveguides to significantly improve these promising results. Ongoing testing is being conducted on water vapor and deuterated water vapor.

- 1 Wilson, E. L., Neveu, M., Riris, H., Georgieva, E. M. & Heaps, W. S. A hollow-waveguide gas correlation radiometer for ultra-precise column measurements of formaldehyde on Mars. *Measurement Science and Technology* **22**, 085902 (2011).
- 2 Drummond, J. R. et al. Early Mission Planning for the MOPITT Instrument. www.atmosp.physics.utoronto.ca/MOPITT/mission.pdf (1999).
- 3 Pfister, G. et al. Effects of a Spectral Surface Reflectance on Measurements of Backscattered Solar Radiation: Application to the MOPITT Methane Retrieval. *Journal of Atmospheric and Oceanic Technology* **22**, 566-574 (2005).
- 4 Tolton, B. T., Hackett, J., Caldwell, D. & Miller, D. Strawman design for a gas-filter correlation radiometer satellite instrument to measure the atmospheric CO₂ column. *Proc. of SPIE* **5543**, 332-337 (2004).

- 5 Tolton, B. T. A Concept for a Gas-Filter Correlation Radiometer to Remotely Sense the Atmospheric Carbon Dioxide Column from Space. *Journal of Atmospheric and Oceanic Technology* **21**, 837-852 (2004).